

DISTRIBUTOR: **AMMEX Corporation**
 ADDRESS: PO BOX 88047, TUKWILA, WA 98138
 DISTRIBUTOR: 1-800-274-7354

PRODUCT NAME(S): GLOVEWORKS POWDERED AND POWDER FREE, TEXTURED LATEX GLOVES

PRODUCT CODES: POWDERED: TL42100, TL44100, TL46100, TL48100

POWDER FREE: TLF42100, TLF44100, TLF46100, TLF48100

SECTION I : PRODUCT DESCRIPTION

Product Name : Latex Examination Glove [Powdered and Powder Free]

Chemical Family : Natural Rubber Latex

SECTION II : COMPOSITION / MAIN RAW MATERIAL

Sulphur, Zinc Oxide, ZDBC, Titanium Dioxide, Calcium Nitrate, Calcium Carbonate, Antioxidant and Corn Starch

PERFORMANCE SPECIFICATION

Dimensions

Description	Size	In-House Standard
Length (mm)	All Sizes	Min 240
Palm Width (mm)	XS	76 +/- 3
	S	84 +/- 3
	M	94 +/- 3
	L	105 +/- 3
	XL	113 +/- 3
Thickness (mm) *single wall	All Sizes	Finger : 0.14 +/- 0.03 Palm : 0.11 +/- 0.03

Physical Properties

Description	In-House Standard	
	Before Aging	After Aging
Elongation at break (%)	min 700	min 500
Tensile Strength (MPa)	min 18	min 14
Median Force at Break (N)	min 9	min 6

The information contained herein is based on the data available to us and is believed to be correct as of the date prepared. However, AMMEX Corporation makes no warranty, expressed or implied regarding the accuracy of these data. For further information, please contact Ammex.

PRODUCT SPECIFICATION Latex Powderfree Glove (Textured)

SECTION I : PRODUCT DESCRIPTION

1.1	Type	Latex Glove, Powderfree , Non-sterile
1.2	Material	Natural High Grade Rubber Latex
1.3	Color	Off white to yellow
1.4	Design and Feature	Ambidextrous, textured , beaded cuff
1.5	Powder	No powder lubricant added
1.6	Storage Condition	The gloves shall maintain their properties when stored in a dry condition at temperature between 10 to 30° C.
1.7	Shelf-Life	The gloves shall have shelf life of 5 years from the date of manufacture with the above storage condition.
1.8	Packing Style	100 pcs gloves x 10 dispensers x 1 carton
1.9	Size Marking	The size of gloves shall be marked in the check box on every carton with black ink.

SECTION II : PERFORMANCE REQUIREMENTS

(Sampling Plan – ISO 2859 Single Normal)

#	Characteristics	Inspection Level	Acceptable Quality Level	Reference Standard
2.1	Dimensions	S2	4.0	ASTM D3578-05
		Median of 13 test pieces		EN455-2-00
2.2	Physical Properties	S2	4.0	ASTM D3578-05
		Median of 13 test pieces		EN455-2-00
2.3	Freedom from Holes (Barrier) (1000ml Water Leak Test)	GI	4.0	ASTM D3578-05 FDA 1000MIL Test Method ASTM D5151-92 EN455-1-00
2.4	Visual Defects:			
(i)	- Major Visual	GI	4.0	In-house practice
(ii)	- Minor Visual		6.5	
2.5	Mix Size / Mix Glove / Mix Hand	Not Allowed		

**Unacceptable at any level

The information contained herein is based on the data available to us and is believed to be correct as of the date prepared. However, AMMEX Corporation makes no warranty, expressed or implied regarding the accuracy of these data. For further information, please contact Ammex.

PRODUCT SPECIFICATION
Latex Powdered Glove (Textured)

SECTION I : PRODUCT DESCRIPTION

1.1	Type	Latex Glove, <u>Powdered</u> , Non-sterile
1.2	Material	Natural High Grade Rubber Latex
1.3	Color (Pigment)	Natural to white and uniform allowing for gradual shift due to thickness variation
1.4	Design and Feature	Ambidextrous, <u>Textured</u> , beaded cuff
1.5	Powder	Inside absorbable cornstarch USPXX1
1.6	Storage Condition	The gloves shall maintain their properties when stored in a dry condition at temperature between 10° to 30° C. Avoid direct sunlight.
1.7	Shelf-Life	The gloves shall have shelf life of 5 years from the date of manufacture with the above storage condition.
1.8	Packing Style	100 pcs gloves x 10 dispensers x 1 carton
1.9	Size Marking	The size of gloves shall be marked in the check box on every carton with black ink.

SECTION II : PERFORMANCE REQUIREMENTS

(Sampling Plan – ISO 2859 Single Normal)

#	Characteristics	Inspection Level	Acceptable Quality Level	Reference Standard
2.1	Dimensions	S2	4.0 -	ASTM D3578-05
			Median of 13 test pieces	EN455-2-00
2.2	Physical Properties	S2	4.0	ASTM D3578-05
			Median of 13 test pieces	EN455-2-00
2.3	Freedom from Holes (Barrier) (1000ml Water Leak Test)	GI	4.0	ASTM D3578-05 FDA 1000MIL Test Method ASTM D5151-92 EN455-1-00
2.4	Visual Defects:			
(i)	- Major Visual	GI	4.0	In-house practice
(ii)	- Minor Visual		6.5	
2.5	Mix Size / Mix Glove / Mix Hand	Not Allowed		

**Unacceptable at any level

The information contained herein is based on the data available to us and is believed to be correct as of the date prepared. However, AMMEX Corporation makes no warranty, expressed or implied regarding the accuracy of these data. For further information, please contact Ammex.